



## CERTIFICATE OF CONFORMANCE

**Product Part Number:** CSLB-136WHIR  
**Lot Number:** 039816-000  
**Irradiation Reference Number:** 0007891-00  
**Lot Quantity:** 6 CASE(S)  
**Date of manufacture:** 09/01/2023\*  
**Shelf Life:** 24 MONTHS  
**Description:** WHITE MICROSEAL TAPE, IRR'D (24/CS)  
**Size -**  
**Width:** 1"  
**Length:** 36 YARDS  
**Height:** N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

9/18/2023

**Authorized Signature**  
**Quality Control Department**

**Date**

**CC:** With Product  
Customer File  
C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

# STERIS: Gamma Certificate Of Processing

Prepared for: MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-16812A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
---------------------	-------------------	-----------------	------------

CSLB-136WHIR	039816-000	6	Case
--------------	------------	---	------

Processing Run Start Date: 17-Sep-2023 1:45 PM

Approx. Downtime (hours): 0.00

Processing Run End Date: 17-Sep-2023 10:37 PM

Minimum Specified Dose (kGy): 25.0

Minimum Delivered Dose (kGy): 27.2

Maximum Specified Dose (kGy): 40.0

Maximum Delivered Dose (kGy): 37.1

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Other Information

0007891-00

## Signature Manifest

Reviewed and E-Signed By: **Luzdaniella Montesinos (Quality Technician)**

Date/Time E-Signed: 18-Sep-2023 10:01 AM

Document Content Revision: 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

## Processing Location

1000 Sarah Place  
Ontario CA 91761  
United States

**APPROVED**

*By Noel Chaidez at 10:05 am, Sep 18, 2023*

Document ID: 170243