



## CERTIFICATE OF CONFORMANCE

**Product Part Number:** PEN-02IR  
**Lot Number:** 049816-000  
**Irradiation Reference Number:** 0010625-00  
**Lot Quantity:** 80 CASE(S)  
**Date of manufacture:** 05/19/2025\*  
**Shelf Life:** 24 MONTHS  
**Description:** CLEANROOM PEN, BLUE, IRR'D (100/CS)  
**Size -**  
**Width:** N/A  
**Length:** N/A  
**Height:** N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

6/2/2025

**Authorized Signature**  
**Quality Control Department**

**Date**

**CC:** With Product  
Customer File  
C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

© COPYRIGHT 2019  
Form No. GEN-1034e  
REV. No. 02

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)  
Gamma Process Run ID 1113-26008A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
PEN-02IR	049816-000	80	Case

Processing Run Start Date	30-May-2025 2:46 PM		
Processing Run End Date	30-May-2025 9:31 PM		
Minimum Specified Dose (kGy)	25.0	Minimum Delivered Dose (kGy)	27.1
Maximum Specified Dose (kGy)	40.0	Maximum Delivered Dose (kGy)	36.9

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information  
0010625-00

**Gamma Process Run Approval authorized by STERIS**  
DateTime Esigned 31-May-2025 4:57 PM

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location  
1000 Sarah Place  
Ontario CA 91761  
United States

Product Code

Lot Number

Quantity   UOM

**APPROVED**  
*By Paola Diaz at 7:10 am, Jun 02, 2025*