



CERTIFICATE OF CONFORMANCE

Product Part Number: CR100-34WHIR
Lot Number: 048658-000
Irradiation Reference Number: 0010625-00
Lot Quantity: 20 CASE(S)
Date of manufacture: 05/21/2025*
Shelf Life: 24 MONTHS
Description: WHITE, VINYL TAPE, IRR'D (48/CS)
Size -
Width: ¾"
Length: 36 YARDS
Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

6/3/2025

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

© COPYRIGHT 2019
Form No. GEN-1034e
REV. No. 02

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)
Gamma Process Run ID 1113-25985A

Product Code	Lot Number	Quantity	UOM
CR100-34WHIR	048658-000	20	Case

Processing Run Start Date 29-May-2025 6:40 AM
Processing Run End Date 29-May-2025 12:06 PM

Minimum Specified Dose (kGy)	27.5	Minimum Delivered Dose (kGy)	30.7
Maximum Specified Dose (kGy)	50.0	Maximum Delivered Dose (kGy)	41.0

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information
0010625-00

Signature Manifest

Reviewed and E-Signed By: Jane Ablola (Quality Analyst)
DateTime Esigned 03-Jun-2025 6:13 AM
Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location
1000 Sarah Place
Ontario CA 91761
United States

APPROVED
By Jenicah Baquir at 8:06 am, Jun 03, 2025