

CERTIFICATE OF CONFORMANCE

Product Part Number:	CR100-1LBIR
Lot Number:	049325-000
Irradiation Reference Number:	0010675-00

Lot Quantity: 3 CASE(S)

Date of manufacture: 06/03/2025*

Shelf Life: 24 MONTHS

Description: TAPE, VINYL, LT BLUE, IRR'D (36/CS)

Size -

Width: 1"

Length: 36 YARDS

Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

6/20/2025

Authorized Signature Quality Control Department Date

CC: With Product Customer File C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

^{*}Refer to Certificate of Processing for irradiation complete date

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-26314A

Product Code Lot Number Quantity UOM

CR100-1LBIR 049325-000 3 Case

Processing Run Start Date 17-Jun-2025 5:29 PM
Processing Run End Date 18-Jun-2025 12:29 AM

Minimum Specified Dose (kGy) 27.5 Minimum Delivered Dose (kGy) 30.3

Maximum Specified Dose (kGy) 50.0 Maximum Delivered Dose (kGy) 40.0

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0010675-00

Signature Manifest

Reviewed and E-Signed By: Tricia Nguyen (Quality Analyst)

DateTime Esigned 20-Jun-2025 8:44 AM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place Ontario CA 91761 United States

APPROVED

By Jenicah Baguir at 8:49 am, Jun 20, 2025

Document ID: 226161

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 1