



## CERTIFICATE OF CONFORMANCE

**Product Part Number:** PEN-40IR  
**Lot Number:** 049811-000  
**Irradiation Reference Number:** 0010701-00  
**Lot Quantity:** 45 CASE(S)  
**Date of manufacture:** 06/11/2025\*  
**Shelf Life:** 24 MONTHS  
**Description:** LABMARKER, BLACK, IPA RESIST., IRR'D (10/PK)  
**Size -**  
**Width:** N/A  
**Length:** N/A  
**Height:** N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

6/20/2025

**Authorized Signature**  
**Quality Control Department**

**Date**

**CC:** With Product  
Customer File  
C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)  
Gamma Process Run ID 1113-26294A

Product Code	Lot Number	Quantity	UOM
PEN-40IR	049811-000	45	Case
Processing Run Start Date 19-Jun-2025 3:19 PM			
Processing Run End Date 19-Jun-2025 5:43 PM			
Minimum Specified Dose (kGy)	10.0	Minimum Delivered Dose (kGy)	14.3
Maximum Specified Dose (kGy)	22.0	Maximum Delivered Dose (kGy)	17.9

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information  
0010701-00

Gamma Process Run Approval authorized by STERIS  
DateTime Esigned 19-Jun-2025 9:00 PM

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location  
1000 Sarah Place  
Ontario CA 91761  
United States

APPROVED  
By Jenicah Baquir at 8:47 am, Jun 20, 2025