

# CERTIFICATE OF CONFORMANCE

Product Part Number: BL1-40TRIRBB

Lot Number: 049798-000

Irradiation Reference Number: 0010701-00

Lot Quantity: 40 CASE(S)

Date of manufacture: 06/10/2025\*

Shelf Life: 36 MONTHS

Description: BINLINER, LDPE, 1 MIL, IRR'D (100/CS)

Size -

Width: 23"

Length: 17

Height: 48"

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

6/19/2025

Authorized Signature
Quality Control Department

Date

CC: With Product Customer File C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

<sup>\*</sup>Refer to Certificate of Processing for irradiation complete date

# **STERIS: Gamma Certificate Of Processing**

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-26324A

Product Code Lot Number Quantity UOM

BL1-40TRIRBB 049798-000 40 Case

Processing Run Start Date 17-Jun-2025 11:51 PM

Processing Run End Date 18-Jun-2025 6:49 AM

Minimum Specified Dose (kGy) 25.0 Minimum Delivered Dose (kGy) 26.2 Maximum Specified Dose (kGy) 40.0 Maximum Delivered Dose (kGy) 36.5

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

#### Other Information

0010701-00

## **Signature Manifest**

Reviewed and E-Signed By: Tricia Nguyen (Quality Analyst)

DateTime Esigned 19-Jun-2025 10:17 AM

**Document Content Revision 1** 

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

### **Processing Location**

1000 Sarah Place Ontario CA 91761 United States

**APPROVED** 

By Jenicah Baguir at 10:21 am, Jun 19, 2025

Document ID: 226100

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 1