

CERTIFICATE OF CONFORMANCE

Product Part Number: BL1-40TRIRBB

Lot Number: 050692-000

Irradiation Reference Number: 0010899-00

Lot Quantity: 40 CASE(S)

Date of manufacture: 07/30/2025*

Shelf Life: 36 MONTHS

Description: BINLINER, LDPE, 1 MIL, IRR'D (100/CS)

Size -

Width: 23"

Length: 17

Height: 48"

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

8/11/2025

Authorized Signature
Quality Control Department

Date

CC: With Product Customer File C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

^{*}Refer to Certificate of Processing for irradiation complete date

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-27080A

Product Code Lot Number Quantity UOM

BL1-40TRIRBB 050692-000 40 Case

Processing Run Start Date 07-Aug-2025 6:59 PM
Processing Run End Date 08-Aug-2025 1:14 AM

Minimum Specified Dose (kGy) 25.0 Minimum Delivered Dose (kGy) 26.7

Maximum Specified Dose (kGy) 40.0 Maximum Delivered Dose (kGy) 37.1

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0010899-00

Signature Manifest

Reviewed and E-Signed By: Jane Ablola (Quality Analyst)

DateTime Esigned 11-Aug-2025 9:54 AM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place Ontario CA 91761 United States

APPROVED

By Paola Diaz at 10:13 am, Aug 11, 2025

Document ID: 230152

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 1