

# CERTIFICATE OF CONFORMANCE

Product Part Number: ITC620-911IR

**Lot Number:** 050151-000

**Irradiation Reference Number:** 0010739-00

Lot Quantity: 2 CASE(S)

Date of manufacture: 06/11/2025\*

Shelf Life: 60 MONTHS

**Description:** ISOLATOR TOOL COVER, MEGATEX/NOVAPOLY, IRR'D

(50/CS)

Size -

Width: 2.5"

Length: 9"

Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

6/30/2025

Authorized Signature
Quality Control Department

Date

CC: With Product Customer File C of C File

<sup>\*</sup>Refer to Certificate of Processing for irradiation complete date

## **STERIS: Gamma Certificate Of Processing**

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-26491A

Product Code Lot Number Quantity UOM

ITC620-911IR 050151-000 2 Case

Processing Run Start Date 28-Jun-2025 4:51 PM
Processing Run End Date 28-Jun-2025 11:33 PM

Minimum Specified Dose (kGy) 25.0 Minimum Delivered Dose (kGy) 28.5

Maximum Specified Dose (kGy) 40.0 Maximum Delivered Dose (kGy) 37.3

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

#### Other Information

0010739-00

### **Signature Manifest**

Reviewed and E-Signed By: Jane Ablola (Quality Analyst)

DateTime Esigned 30-Jun-2025 7:23 AM

**Document Content Revision 1** 

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

#### **Processing Location**

1000 Sarah Place Ontario CA 91761 United States

Document ID: 226866

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 2

**APPROVED** 

By Jenicah Baquir at 8:39 am, Jun 30, 2025

Document ID: 226866

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 2 of 2