

CERTIFICATE OF CONFORMANCE

Product Part Number: BL2-40TRIRB

Lot Number: 051333-000

Irradiation Reference Number: 0010991-00

Lot Quantity: 192 CASE(S)

Date of manufacture: 08/19/2025*

Shelf Life: 36 MONTHS

Description: BINLINER, LDPE, 2 MIL, BULK, IRR'D (50/CS)

Size -

Width: 40"

Length: 48"

Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

8/26/2025

Authorized Signature
Quality Control Department

Date

CC: With Product Customer File C of C File

*Refer to Certificate of Processing for irradiation complete date

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-27343A

Product Code Lot Number Quantity UOM

BL2-40TRIRB 051333-000 192 Case

Processing Run Start Date 25-Aug-2025 2:13 PM
Processing Run End Date 25-Aug-2025 10:11 PM

Minimum Specified Dose (kGy) 25.0 Minimum Delivered Dose (kGy) 26.6

Maximum Specified Dose (kGy) 40.0 Maximum Delivered Dose (kGy) 37.6

PO# 0010991-00

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Gamma Process Run Approval authorized by STERIS

DateTime Esigned 26-Aug-2025 1:44 AM

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place Ontario CA 91761 United States

APPROVED

By Jenicah Baquir at 10:22 am, Aug 26, 2025

Document ID: 232332

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 1