

# CERTIFICATE OF CONFORMANCE

Product Part Number:	MZSM56-9IR
Lot Number:	049237-000

Irradiation Reference Number: 0010557-00

Lot Quantity: 24 CASE(S)

Date of manufacture: 05/07/2025\*

Shelf Life: 60 MONTHS

**Description:** MZONE SNAPMOP NOVAPOLY, IRR'D (24/CS)

Size -

Width: 5"

Length: 9"

Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

5/16/2025

Authorized Signature
Quality Control Department

Date

CC: With Product Customer File C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

<sup>\*</sup>Refer to Certificate of Processing for irradiation complete date

# **STERIS: Gamma Certificate Of Processing**

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-25766A

Product Code Lot Number Quantity UOM

MZSM56-9IR 049237-000 24 Case

Processing Run Start Date 13-May-2025 4:47 AM

Processing Run End Date 13-May-2025 11:12 AM

Minimum Specified Dose (kGy) 25.0 Minimum Delivered Dose (kGy) 28.9

Maximum Specified Dose (kGy) 40.0 Maximum Delivered Dose (kGy) 38.4

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

#### **Other Information**

0010557-00

## **Signature Manifest**

Reviewed and E-Signed By: Jane Ablola (Quality Analyst)

DateTime Esigned 16-May-2025 7:16 AM

**Document Content Revision 1** 

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

### **Processing Location**

1000 Sarah Place Ontario CA 91761 United States

**APPROVED** 

By Jenicah Baquir at 7:41 am, May 16, 2025

Document ID: 223434

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 1