



## CERTIFICATE OF CONFORMANCE

**Product Part Number:** CSLB-2WHIR  
**Lot Number:** 048319-000  
**Irradiation Reference Number:** 0010675-00  
**Lot Quantity:** 6 CASE(S)  
**Date of manufacture:** 04/30/2025\*  
**Shelf Life:** 24 MONTHS  
**Description:** WHITE MICROSEAL TAPE, IRR'D (24/CS)  
**Size -**  
**Width:** 2"  
**Length:** 18 YARDS  
**Height:** N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

6/20/2025

**Authorized Signature**  
**Quality Control Department**

**Date**

**CC:** With Product  
Customer File  
C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)  
Gamma Process Run ID 1113-26314A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
CSLB-2WHIR	048319-000	6	Case

Processing Run Start Date	17-Jun-2025 5:29 PM		
Processing Run End Date	18-Jun-2025 12:29 AM		
Minimum Specified Dose (kGy)	27.5	Minimum Delivered Dose (kGy)	30.3
Maximum Specified Dose (kGy)	50.0	Maximum Delivered Dose (kGy)	40.0

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information  
0010675-00

Signature Manifest

Reviewed and E-Signed By: **Tricia Nguyen (Quality Analyst)**  
DateTime Esigned 20-Jun-2025 8:44 AM  
  
Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location  
1000 Sarah Place  
Ontario CA 91761  
United States

**APPROVED**  
*By Jenicah Baquir at 8:49 am, Jun 20, 2025*