



## CERTIFICATE OF CONFORMANCE

<b>Product Part Number:</b>	PEN-22IR
<b>Lot Number:</b>	051168-000
<b>Irradiation Reference Number:</b>	0011123-00
<b>Lot Quantity:</b>	11 CASE(S)
<b>Date of manufacture:</b>	09/12/2025*
<b>Shelf Life:</b>	24 MONTHS
<b>Description:</b>	SHARPIE, FINE POINT, BLUE, IRR'D (12/PK)
<b>Size -</b>	
<b>Width:</b>	N/A
<b>Length:</b>	N/A
<b>Height:</b>	N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

9/30/2025

---

**Authorized Signature**  
**Quality Control Department**

**Date**

CC:    With Product  
         Customer File  
         C of C File

\*Refer to Certificate of Processing for irradiation complete date

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

© COPYRIGHT 2019  
Form No. GEN-1034e  
REV. No. 02

# STERIS: Gamma Certificate Of Processing

Prepared For **MICRONOVA MFG INC (2727)**

Gamma Process Run ID **1113-27723A**

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
<b>PEN-22IR</b>	<b>051168-000</b>	<b>11</b>	<b>Case</b>

---

Processing Run Start Date **25-Sep-2025 10:23 AM**

Processing Run End Date **25-Sep-2025 9:56 PM**

<b>Minimum Specified Dose (kGy)</b>	<b>25.0</b>	<b>Minimum Delivered Dose (kGy)</b>	<b>28.9</b>
<b>Maximum Specified Dose (kGy)</b>	<b>40.0</b>	<b>Maximum Delivered Dose (kGy)</b>	<b>39.3</b>

PO# 0011123-00

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Signature Manifest

Reviewed and E-Signed By: **Tricia Nguyen (Quality Analyst)**

DateTime Esigned 30-Sep-2025 6:47 AM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

## Processing Location

**1000 Sarah Place  
Ontario CA 91761  
United States**

Document ID: 234778

Product Code

Lot Number

Quantity    UOM

**APPROVED**

*By Jenicah Baquir at 2:14 pm, Sep 30, 2025*

Document ID: 234778