



## CERTIFICATE OF CONFORMANCE

<b>Product Part Number:</b>	PEN-22IR
<b>Lot Number:</b>	052926-000
<b>Irradiation Reference Number:</b>	0011350-00
<b>Date of manufacture:</b>	11/06/2025
<b>Manufacturing/Processing Site:</b>	MIRCONOVA MANUFACTURING LLC 3431 WEST LOMITA BLVD. TORRANCE, CA 90505
<b>Irradiation Site:</b>	STERIS ISOMEDIX 1000 SARAH PL. ONTARIO, CA 91761
<b>Shelf Life:</b>	24 MONTHS*
<b>Description:</b>	SHARPIE, FINE POINT, BLUE, IRR'D (12/PK)
<b>Size -</b>	
<b>Width:</b>	N/A
<b>Length:</b>	N/A
<b>Height:</b>	N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

11/17/2025

---

**Authorized Signature**  
**Quality Control Department**

**Date**

CC:    With Product  
         Customer File  
         C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 • 301 LEORA LANE (BUILDING 4, SUITE 420), THE COLONY, TX 75056  
Tel. (310) 784-6990 • FAX (310) 784-6980

# STERIS: Gamma Certificate Of Processing

Prepared For **MICRONOVA MFG INC (2727)**

Gamma Process Run ID **1113-28431A**

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
<b>PEN-22IR</b>	<b>052926-000</b>	<b>15</b>	<b>Case</b>

Processing Run Start Date **11-Nov-2025 4:37 PM**

Processing Run End Date **11-Nov-2025 10:37 PM**

<b>Minimum Specified Dose (kGy)</b>	<b>25.0</b>	<b>Minimum Delivered Dose (kGy)</b>	<b>27.6</b>
<b>Maximum Specified Dose (kGy)</b>	<b>40.0</b>	<b>Maximum Delivered Dose (kGy)</b>	<b>36.6</b>

PO# 0011350-00

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Signature Manifest

Reviewed and E-Signed By: **Jane Ablola (Quality Analyst)**

DateTime Esigned 13-Nov-2025 6:29 AM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

## Processing Location

**1000 Sarah Place  
Ontario CA 91761  
United States**

Document ID: 238104

Product Code

Lot Number

Quantity    UOM

**APPROVED**

*By Jenicah Baquir at 8:13 am, Nov 13, 2025*

Document ID: 238104