



CERTIFICATE OF CONFORMANCE

| | |
|--------------------------------------|--|
| Product Part Number: | PEN-20IRDB |
| Lot Number: | 048122-000 |
| Irradiation Reference Number: | 0010675-00 |
| Lot Quantity: | 6 CASE(S) |
| Date of manufacture: | 06/04/2025* |
| Shelf Life: | 24 MONTHS |
| Description: | SHARPIE FINE, BLACK, EA. DBL BAGGED, IRR'D (12/PK) |
| Size - | |
| Width: | N/A |
| Length: | N/A |
| Height: | N/A |

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

6/23/2025

Authorized Signature
Quality Control Department

Date

CC: With Product
 Customer File
 C of C File

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared For **MICRONOVA MFG INC (2727)**

Gamma Process Run ID **1113-26359A**

| <u>Product Code</u> | <u>Lot Number</u> | <u>Quantity</u> | <u>UOM</u> |
|---------------------|-------------------|-----------------|-------------|
| PEN-20IRDB | 048122-000 | 6 | Case |

Processing Run Start Date **19-Jun-2025 9:22 PM**

Processing Run End Date **20-Jun-2025 2:44 AM**

| | | | |
|-------------------------------------|-------------|-------------------------------------|-------------|
| Minimum Specified Dose (kGy) | 25.0 | Minimum Delivered Dose (kGy) | 27.1 |
| Maximum Specified Dose (kGy) | 40.0 | Maximum Delivered Dose (kGy) | 35.9 |

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0010675-00

Signature Manifest

Reviewed and E-Signed By: **Jane Ablola (Quality Analyst)**

DateTime Esigned 23-Jun-2025 9:56 AM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

**1000 Sarah Place
Ontario CA 91761
United States**

APPROVED

By Jenicah Baquir at 10:04 am, Jun 23, 2025

Document ID: 226298