



## CERTIFICATE OF CONFORMANCE

<b>Product Part Number:</b>	PEN-20IRDB
<b>Lot Number:</b>	050993-000
<b>Irradiation Reference Number:</b>	0011006-00
<b>Lot Quantity:</b>	6 CASE(S)
<b>Date of manufacture:</b>	08/22/2025*
<b>Shelf Life:</b>	24 MONTHS
<b>Description:</b>	SHARPIE FINE, BLACK, EA. DBL BAGGED, IRR'D (12/PK)
<b>Size -</b>	
<b>Width:</b>	N/A
<b>Length:</b>	N/A
<b>Height:</b>	N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

9/8/2025

---

**Authorized Signature**  
**Quality Control Department**

**Date**

CC:    With Product  
         Customer File  
         C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

# STERIS: Gamma Certificate Of Processing

Prepared For **MICRONOVA MFG INC (2727)**

Gamma Process Run ID **1113-27455A**

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
<b>PEN-20IRDB</b>	<b>050993-000</b>	<b>6</b>	<b>Case</b>

Processing Run Start Date **04-Sep-2025 3:33 AM**

Processing Run End Date **04-Sep-2025 9:14 AM**

**Minimum Specified Dose (kGy)** **25.0**      **Minimum Delivered Dose (kGy)** **27.1**

**Maximum Specified Dose (kGy)** **40.0**      **Maximum Delivered Dose (kGy)** **34.4**

PO# 0011006-00

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Signature Manifest

Reviewed and E-Signed By: **Tricia Nguyen (Quality Analyst)**

DateTime Esigned 08-Sep-2025 11:54 AM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

## Processing Location

**1000 Sarah Place  
Ontario CA 91761  
United States**

**APPROVED**

*By Jenicah Baquir at 12:50 pm, Sep 08, 2025*

Document ID: 233337