



CERTIFICATE OF CONFORMANCE

Product Part Number: BL2-32YEIRBB
Lot Number: 051676-000
Irradiation Reference Number: 0011050-00
Lot Quantity: 20 CASE(S)
Date of manufacture: 09/04/2025*
Shelf Life: 36 MONTHS
Description: BINLINER, YELLOW BIOHAZARD, IRR'D (50/CS)
Size -
Width: 32"
Length: 36"
Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

9/16/2025

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

***Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)
Gamma Process Run ID 1113-27578A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
BL2-32YEIRBB	051676-000	20	Case

Processing Run Start Date 11-Sep-2025 5:35 AM

Processing Run End Date 11-Sep-2025 1:53 PM

Minimum Specified Dose (kGy)	25.0	Minimum Delivered Dose (kGy)	26.3
Maximum Specified Dose (kGy)	40.0	Maximum Delivered Dose (kGy)	36.5

PO# 0011050-00

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By: Julie Candelaria-Guerrero (Quality Analyst)

DateTime Esigned 15-Sep-2025 2:58 PM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place
Ontario CA 91761
United States

APPROVED

By Jenicah Baquir at 3:00 pm, Sep 15, 2025