



## CERTIFICATE OF CONFORMANCE

**Product Part Number:** BL3-2436IRBB  
**Lot Number:** 049992-000  
**Irradiation Reference Number:** 0010739-00  
**Lot Quantity:** 60 CASE(S)  
**Date of manufacture:** 06/23/2025\*  
**Shelf Life:** 36 MONTHS  
**Description:** BINLINER, LDPE, 3 MIL, IRR'D (50/CS)  
**Size -**  
**Width:** 24"  
**Length:** 36"  
**Height:** N/A

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

6/30/2025

**Authorized Signature**  
**Quality Control Department**

**Date**

**CC:** With Product  
Customer File  
C of C File

**\*Refer to Certificate of Processing for irradiation complete date**

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

© COPYRIGHT 2019  
Form No. GEN-1034e  
REV. No. 02

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)  
Gamma Process Run ID 1113-26491A

<u>Product Code</u>	<u>Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
BL3-2436IRBB	049992-000	60	Case

Processing Run Start Date 28-Jun-2025 4:51 PM  
Processing Run End Date 28-Jun-2025 11:33 PM

Minimum Specified Dose (kGy)	25.0	Minimum Delivered Dose (kGy)	28.5
Maximum Specified Dose (kGy)	40.0	Maximum Delivered Dose (kGy)	37.3

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information  
0010739-00

Signature Manifest

Reviewed and E-Signed By: Jane Ablola (Quality Analyst)  
DateTime Esigned 30-Jun-2025 7:23 AM  
Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location  
1000 Sarah Place  
Ontario CA 91761  
United States

Product Code

Lot Number

Quantity   UOM

**APPROVED**

*By Jenicah Baquir at 8:39 am, Jun 30, 2025*