

CERTIFICATE OF CONFORMANCE

Product Part Number: PEN-52IR

Lot Number: 048796-000

Irradiation Reference Number: 0010690-00

Lot Quantity: 5 CASE(S)

Date of manufacture: 06/06/2025*

Shelf Life: 24 MONTHS

Description: SHARPIE ULTRA FINE, BLUE, IRR'D (12/PK)

Size -

Width: N/A

Length: N/A

Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

6/20/2025

Authorized Signature
Quality Control Department

Date

CC: With Product Customer File C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

^{*}Refer to Certificate of Processing for irradiation complete date

STERIS: Gamma Certificate Of Processing

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-26248A

Product Code Lot Number Quantity UOM

PEN-52IR	048796-000	5	Case
Processing Run Start Date 12-Jun-2025 8:3	8 PM		
Processing Run End Date 13-Jun-2025 1:2	2 AM		
Minimum Specified Dose (kGy) 2	5.0 Minimum Delivered Dose (kGy)	27.8
Maximum Specified Dose (kGy) 4	0.0 Maximum Delivered Dose ((kGy)	37.2

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Other Information

0010690-00

Signature Manifest

Reviewed and E-Signed By: **Bharat Rupapara (Quality Analyst)**

DateTime Esigned 20-Jun-2025 9:54 AM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

Processing Location

1000 Sarah Place Ontario CA 91761 United States

APPROVED

By Jenicah Baquir at 10:18 am, Jun 20, 2025

Document ID: 226171

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 1