

# CERTIFICATE OF CONFORMANCE

Product Part Number:	BL4-3440IRBB

**Lot Number:** 050572-000

**Irradiation Reference Number:** 0010810-00

Lot Quantity: 128 CASE(S)

Date of manufacture: 07/09/2025\*

Shelf Life: 36 MONTHS

**Description:**BINLINER, LDPE, 4 MIL, IRR'D (50/CS)

Size -

Width: 34"

Length: 40"

Height: N/A

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

7/25/2025

Authorized Signature
Quality Control Department

Date

CC: With Product Customer File C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6980

<sup>\*</sup>Refer to Certificate of Processing for irradiation complete date

## **STERIS: Gamma Certificate Of Processing**

Prepared For MICRONOVA MFG INC (2727)

Gamma Process Run ID 1113-26807A

 Product Code
 Lot Number
 Quantity
 UOM

 BL4-3440IRBB
 050572-000
 128
 Case

Processing Run Start Date 23-Jul-2025 2:47 AM

Processing Run End Date 23-Jul-2025 8:00 AM

Minimum Specified Dose (kGy) 25.0 Minimum Delivered Dose (kGy) 28.3

Maximum Specified Dose (kGy) 40.0 Maximum Delivered Dose (kGy) 37.7

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

#### Other Information

0010810-00

### **Signature Manifest**

Reviewed and E-Signed By: **Bharat Rupapara (Quality Analyst)** 

DateTime Esigned 25-Jul-2025 2:12 PM

Document Content Revision 1

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2016, and in alignment with EN ANSI/AAMI/ISO 11137:2017. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.

#### **Processing Location**

1000 Sarah Place Ontario CA 91761 United States

**APPROVED** 

By Jenicah Baquir at 2:52 pm, Jul 25, 2025

Document ID: 228827

GWI-08-006-15 Last Revised in Rel 2.0.0.0 Rel Date: 13-Nov-2018 Page 1 of 1