



# Certificate of Compliance

**Product description:** Shaker Flask, 3000 ml flat bottom

**Greiner part no.:** 679506  
**Expiration date:** May 27, 2027

**Greiner lot no.:** 1 1 0 4 2 7 - -

**Physical inspection:**  passed

**Visual inspection:**  passed

**Product has been gamma irradiated** and dosimetrically released based upon U.S. Association for the Advancement of Medical Instrumentation (AAMI) Recommended practices. Gamma sterilized to SAL 10<sup>-6</sup>

The material listed has received the following doses within the precision limits of the dosimetry system employed

Minimum Dosage: 29.1 kGy Maximum Dosage: 36.2 kGy

Isotope Utilized: 60 COBALT-60

Dose Requirement: Min: 18kGy; Max: 30kGy

**Pyrogens** An extract from the lot contained < 0.005 EU/ml per the Limulus Amoebocyte Lysate (LAL) test method in the *FDA Guidelines on Validation of the Limulus Amoebocyte Lysate Test as an end product Endotoxin test for Human and Animal Parenteral Drugs, Biological Products, and Medical Devices*, December 1987.

**RNase-free, DNase-free, DNA-free and Non-pyrogenic.**

**We hereby certify that this shipment is in accordance with the valid specifications.**

31 Aug 2022  
Date

Vickie Kilpatrick  
Quality Assurance

Greiner Bio-One NA Inc  
4238 Capital Drive  
Monroe, NC 28110, USA