



# Certificate of Analysis

### pH Buffer Solution

## Buffer Solution pH 10.00 ± 0.01 @ 20°C

 Product No:
 1100
 Date of Measurement:
 11/12/2024

 Lot No:
 1024M1
 Date of Sample Receipt:
 11/12/2024

 Expiry Date\*:
 28/12/2026
 Date of Manufacture\*:
 11/12/2024

Specification: Mean Measured Value:

9.99 - 10.01 @ 20°C 10.01 pH @ 20°C

#### Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPPHB. Measured with a combination glass electrode after multiple point calibration with reference materials. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above. The uncertainty of measurement has been calculated not to exceed ± 0.010pH at 95% confidence level, i.e. coverage factor k =2.

#### **Buffer Substance:**

Di-sodium Tetraborate, Sodium Hydroxide.

#### Metrological Traceability:

The test result is traceable to Standard Reference Material of National Institute of Standards and Technology (USA), SRM 191d-I Sodium Bicarbonate, SRM 191d-II Sodium Carbonate, SRM 186-I-g Potassium Dihydrogen Phosphate and SRM 186-II-g Disodium Hydrogen Phosphate.

#### Compliance:

This Test Method is in accordance with IUPAC Recommendations 2002 Measurement of pH. Definition, Standards and Procedures.

#### Accreditation:

Reagecon Diagnostics Ltd. is accredited to ISO 17025 by the American Association for Laboratory Accreditation, under scope 6739.03, for the test method, TPPHB, used to generate the above result. This accreditation deems Reagecon competent on a quality systems level and a technical level to perform the tests on the scope of accreditation. Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPPHB is technically valid and is supported by appropriate uncertainty measurements.

Date of Issue of the Certificate:

12/12/2024

QC Laboratory Techinician

**ONDIANO** Gabriel

gabriel