

CERTIFICATE OF GRAVIMETRIC PREPARATION

PRODUCT: Concentrate to make Nitrate Standard in H₂O 5002100
in accordance with European Pharmacopoeia

PRODUCT No.: 5002100C

LOT NO.: 521C25F1

DATE OF PREPARATION: 30th June 2025

EXPIRY DATE: 28th June 2026

DENSITY VALUE: 0.999 g/ml @ 20°C

PREPARATION OF CONCENTRATE:

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

| Analyte | Raw Material | Nominal mg/kg | Actual mg/kg |
|-----------------------------|-------------------|---------------|--------------|
| Nitrate, as NO ₃ | Potassium nitrate | 1000 | 1001 ± 0.2 % |

1001 mg/kg is equivalent to 1000 mg/L @ 20°C

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as ± 0.2 %. All values are verified by IC analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

DILUTION INSTRUCTIONS FOR PREPARATION OF Ph. Eur. 5002100 (100ppm NO₃ aq.)

1. To prepare Ph. Eur. 5002100 (100ppm Nitrate, as NO₃) dilute this solution to 10 times it's volume with purified water.
2. Prepare the dilute solution immediately before use.

TRACEABILITY IN THE PRODUCTION OF THIS STANDARD:

This product was prepared gravimetrically on a mass/mass basis, using a balance calibrated by Reagecon engineers with mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines. The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity, and linearity.

TEST METHOD:

The mean result of this standard was verified using a calibrated IC system according to an in-house test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high-performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By: QC Supervisor



Date: 16th July 2025

This certificate must not be reproduced except in full.