

Certificate of Analysis

European Pharmacopoeia

0.1 M Tetrabutyl Ammonium Hydroxide in 2-Propanol

Product No:	3008400	Date of Measurement:	30/01/2025
Lot No:	384025A1	Date of Sample Receipt:	30/01/2025
Expiry Date*:	28/01/2026	Date of Manufacture*:	30/01/2025

Specification:

0.0997 - 0.1003M

Mean Measured Value:

0.0997 M @ 20°C

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. This Solution was checked by means of Perchloric Acid in Acetic Acid Analytical Volumetric Solution. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above.

Metrological Traceability:

The test result is directly traceable to Standard Reference Material of National Institute of Standards and Technology (USA), SRM 84L Potassium Hydrogen Phthalate.


Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :

31/01/2025

QC Laboratory Technician

ONDIANO Gabriel

A handwritten signature in black ink that reads "gabriel".

All raw materials used to prepare this product are of high purity.

*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL001