

Certificate of Analysis

European Pharmacopoeia

Potassium Hydroxide 0.1M (0.1N)

Product No:
Lot No:
Expiry date:

3004800
KOH20121B1
28/02/2023

Mean Molarity:
Date of measurement:
Specification:
0.0997 - 0.1003M

0.0999 M @ 20°C
02/02/2021

Method:

Tested by potentiometric titration, to in-house method TPATB.

Traceability:

This volumetric solution was checked by means of Hydrochloric acid Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 723e Tris(hydroxymethyl) aminomethane. All raw materials used to prepare this product are of high purity.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

Date of issue of the certificate: 08/04/2021



QC Technician **Patricia de Guzman**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21C65

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com