

## Certificate of Analysis

European Pharmacopoeia

Potassium Hydroxide 0.1M (0.1N)

Product No:	3004800
Lot No:	KOH20121K1
Expiry date * :	28/10/2023

Mean Molarity:	0.1000 M @ 20°C
Date of measurement:	06/10/2021
Date of Sample Receipt:	06/10/2021
Date of Manufacturing * :	06/10/2021

\* The detail above is based on information supplied in writing by Reagecon Manufacturing.

**Specification:**

0.0997 - 0.1003M

**Method:**

Tested by potentiometric titration, to in-house method TPATB.

**Traceability:**

This volumetric solution was checked by means of Hydrochloric acid Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 723e Tris(hydroxymethyl) aminomethane. All raw materials used to prepare this product are of high purity.

Date of issue of the certificate: 06/10/2021

*Eileen Malone*

QC Technician **Eileen Malone**

Complementary information relative to this product is available at [www.reagecon.com](http://www.reagecon.com)  
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21K3

Tested by Reagecon Quality Control  
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.  
Shannon Free Zone, Shannon, Co. Clare, Ireland.  
Tel +353 61 472622, Fax: +353 61 472642  
Email: [sales@reagecon.ie](mailto:sales@reagecon.ie), [www.reagecon.com](http://www.reagecon.com)