

Shannon Free Zone, Shannon, Co. Clare, Ireland Tel: +353 61 472622 Fax: +353 61 472642 Email:sales@reagecon.ie

www.reagecon.com

# **CERTIFICATE OF GRAVIMETRIC PREPARATION**

**PRODUCT:** ICP Standard Sulphur 1000 μg/mL

**PRODUCT No.:** PS2A7

**MATRIX:**  $H_2O$ 

**LOT NO.:** PS2724K1

**DATE OF PREPARATION:** 08<sup>th</sup> October 2024

**EXPIRY DATE:** 28<sup>th</sup> October 2026

**DENSITY VALUE:** 1.001 g/ml @ 20 °C

### PREPARATION OF STANDARD:

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

Raw Material	Purity %	Nominal µg/mL	Actual mg/kg
Ammonium Sulphate	99.999	1000	999

### 999 mg/kg is equivalent to 1000 mg/L @ 20 °C

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as  $\pm$  0.2 %. All values are verified by ICP-MS analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

### TRACEABILITY IN THE PRODUCTION OF THIS STANDARD:

This product was prepared gravimetrically on a mass/mass basis, using balances calibrated by Reagecon engineers with mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines. The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity, and linearity.

## **TEST METHOD:**

The mean result of this standard was verified using a calibrated ICP-MS system according to an in-house test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high-performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By: QC Technician

**Date:** 12<sup>th</sup> November 2024

gabriel

This certificate must not be reproduced except in full.