

Shannon Free Zone, Shannon, Co. Clare, Ireland Tel: +353 61 472622 Fax: +353 61 472642 Email:sales@reagecon.ie <u>www.reagecon.com</u>

# **CERTIFICATE OF GRAVIMETRIC PREPARATION**

<b>PRODUCT:</b>	ICP Standard Yttrium 1000 mg/L in 2-5% HNO3	
PRODUCT No.:	PY2A2	
LOT NO.:	PY2225D1	
DATE OF PREPARATION:	1st April 2025	
EXPIRY DATE:	28 <sup>th</sup> April 2027	
DENSITY VALUE:	1.015 g/mL @ 20°C	

## **PREPARATION OF STANDARD:**

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

Raw Material	Purity %	Nominal mg/kg	Actual mg/kg
Yttrium Oxide	99.999	982	$987\pm0.2~\%$

#### 987 mg/kg is equivalent to 1003 mg/L @ 20°C

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as  $\pm 0.2$  %. All values are verified by ICP-MS analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

## TRACEABILITY IN THE PRODUCTION OF THIS STANDARD

This product was prepared gravimetrically on a mass/mass basis, using a balance calibrated by Reagecon engineers with mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines. The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity and linearity.

### **TEST METHOD:**

The mean result of this standard was verified using a calibrated ICP-MS system according to an in-house test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By: QC Technician

Jamillin

Date: 14<sup>th</sup> April 2025

This certificate must not be reproduced except in full.