

# Certificate of Conformity

## Indicator Solution

### Universal Indicator for pH Determination of Aqueous Solutions (pH4 - 10)

Product No:	UN105	Date of Measurement:	16/06/2022
Lot No:	UN122F1	Date of Sample Receipt:	16/06/2022
Expiry Date*:	28/06/2024	Date of Manufacture*:	16/06/2022

#### Sensitivity Test:

Colour Change: pH 4 Pink / pH 5 Orange / pH 6 Yellow / pH 7 Green / pH 8 Turquoise / pH 9 Blue / pH 10 Purple.

Compliant

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :  
20/06/2022

QA Officer  
COONEY Rosemary

*Rosemary Cooney.*

All raw materials used to prepare this product are of high purity.  
\*The detail above is based on information supplied in writing by Reagecon Manufacturing.  
Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing  
This Certificate must not be reproduced except in full. Rev-QL001