

Certificate of Analysis

Turbidimetric Calibration Standard

0.2 NTU Non Ratio Turbidity Standard

Product No:	CRS-0.2-100	Date of Measurement:	23/05/2025
Lot No:	CRS0225E1	Date of Sample Receipt:	23/05/2025
Expiry Date*:	28/05/2027	Date of Manufacture*:	23/05/2025

Specification:	Mean Measured Value:
0.18 - 0.22 NTU	0.19 NTU

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above.

Metrological Traceability:

The test result is directly traceable to the Formazin Primary Standard for Turbidity, as adopted by NIST and other National Metrology Institutes.

Reference:

Water Quality - Determination of Turbidity - International Standard ISO 7027 and ASTM D 1889.

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :
23/05/2025

QC Laboratory Technician
PILI Jashil

A handwritten signature in black ink, appearing to read "PILI Jashil".

All raw materials used to prepare this product are of high purity.
*The detail above is based on information supplied in writing by Reagecon Manufacturing.
Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing
This Certificate must not be reproduced except in full. Rev-QL003