

Certificate of Analysis

Electrode Care and Maintenance

1 mol/l Lithium Chloride, dissolved in Isopropanol

Product No:
Lot No:
Expiry Date:

EFSLICL
EFSLC20J1
28/09/2023

Mean Measured Value:
Date of Measurement:

1.0 M
18/09/2020

Specification:

0.9 M - 1.1 M Lithium Chloride

Method:

Tested via Cl in accordance with In-House Argentometric Titration Test Method.

Traceability:

This solution was checked by means of Silver Nitrate Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), SRM 919b Sodium Chloride.

Reference:

Vogels Textbook of Quantitative Inorganic Analysis.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

Date of Issue of the Certificate: 18/09/2020

Eileen Malone

QC Technician Eileen Malone

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-20J9

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com