

Certificate of Analysis

Analytical Volumetric Solution

Sodium Nitrite 0.5M (0.5N)

Product No:
Lot No:
Expiry date * :

NANO051
NANO0521M1
28/12/2022

Mean measured value:
Date of measurement:
Date of Sample Receipt:
Date of Manufacturing * :

0.4977 M @ 20°C
14/12/2021
14/12/2021
14/12/2021

* The detail above is based on information supplied in writing by Reagecon Manufacturing.

Specification:

0.4975 - 0.5025M

Method:

Tested by potentiometric titration, to in-house method.

Traceability:

This volumetric solution was checked by means of Sodium Thiosulphate Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 136f Potassium Dichromate. All raw materials used to prepare this product are of high purity.

Date of issue of the certificate:

14/12/2021

Eileen Malone

QC Technician **Eileen Malone**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21M7

Tested by Reagecon Quality Control
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com