

Certificate of Analysis

Standards and Reagents for Chloride Analysers

Combined Acid Buffer

| | | | |
|---------------|------------|-------------------------|------------|
| Product No: | CAB05 | Date of Measurement: | 02/03/2023 |
| Lot No: | CAB0523C1 | Date of Sample Receipt: | 02/03/2023 |
| Expiry Date*: | 28/03/2025 | Date of Manufacture*: | 02/03/2023 |

Specification:

Total Acidity: 1.8 - 1.9 M

Mean Measured Value:

1.8 M @ 20°C

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Tested by Potentiometric Titration to In-House Test Method. This Solution was checked by means of Sodium Hydroxide Analytical Volumetric Standard. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above.

Metrological Traceability:

The test result is directly traceable to Standard Reference Material of National Institute of Standards and Technology (USA), SRM 84L Potassium Hydrogen Phthalate.


Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :

03/03/2023

Senior QC Technician

MALONE Eileen

A handwritten signature in black ink that reads "Eileen Malone".

All raw materials used to prepare this product are of high purity.

*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL001