



# Certificate of Analysis

## Electrode Care and Maintenance

### Electrode Filling Solution 3.0M KCl / AgCl

*Product No:*

EFS3AC

*Lot No:*

EFS3AC20E1

*Expiry Date:*

28/05/2022

*Mean Measured Value:*

3.0 M

*Date of Measurement:*

14/05/2020

**Specification:**

2.9 - 3.1 M KCl

#### **Method:**

Tested via Cl in accordance with In-House Argentometric Titration Test Method.

#### **Traceability:**

This solution was checked by means of Silver Nitrate Analytical Volumetric Standard. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 919b Sodium Chloride. All raw materials used to prepare this product are of high purity.

#### **Reference:**

Vogels Textbook of Quantitative Inorganic Analysis.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

*Date of Issue of the Certificate:*

14/05/2020

*Eileen Malone*

**QC Technician Eileen Malone**

Complementary information relative to this product is available at [www.reagecon.com](http://www.reagecon.com)  
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-20E8

Reagecon Diagnostics Ltd.  
Shannon Free Zone, Shannon, Co. Clare, Ireland.  
Tel +353 61 472622, Fax: +353 61 472642  
Email: [sales@reagecon.ie](mailto:sales@reagecon.ie), [www.reagecon.com](http://www.reagecon.com)