

Shannon Free Zone, Shannon, Co. Clare, Ireland Tel: +353 61 472622 Fax: +353 61 472642 Email:sales@reagecon.ie

www.reagecon.com

CERTIFICATE OF GRAVIMETRIC PREPARATION

PRODUCT:	IC Standard Iron 1000 mg/L	
PRODUCT No.:	ICCS12	
MATRIX:	$H_2O + tr HNO_3$	
LOT NO.:	ICCS1222K1	
DATE OF PREPARATION:	04 th April 2023	
EXPIRY DATE:	28 th October 2024	
DENSITY VALUE:	1.001 g/ml @ 20°C	

PREPARATION OF STANDARD:

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

Analyte	Raw Material	Nominal mg/kg	Actual mg/kg
Iron, as Fe	Iron Nitrate	999	$999\pm0.2~\%$

999 mg/kg is equivalent to 1000 mg/L @ 20°C

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as ± 0.2 %. All values are verified by ICP-MS analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

TRACEABILITY IN THE PRODUCTION OF THIS STANDARD

This product was prepared gravimetrically on a mass/mass basis. The solute was weighed on a balance calibrated by Reagecon engineers using mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines. The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity, and linearity.

TEST METHOD:

The mean result of this standard was verified using a calibrated ICP-MS system according to an in-house test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high-performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By:

Eleen Malone.

Date: 04th April 2023

This certificate must not be reproduced except in full.