



Certificate of Analysis

Conductivity Standard Solution

350000 µS/cm @ 25°C

 Product No:
 CSKC350M
 Date of Measurement:
 14/11/2024

 Lot No:
 CS350M24L1
 Date of Sample Receipt:
 14/11/2024

 Expiry Date*:
 28/05/2026
 Date of Manufacture*:
 14/11/2024

 Specification:
 Mean Measured Value:

 346500 - 353500 μS/cm @ 25°C
 348801 μS/cm @ 25°C

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPCOND. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above. The uncertainty of measurement has been calculated not to exceed ± 1% at 95% confidence level, k=2.

Metrological Traceability:

Measurement taken by comparison with standard prepared from National Institute of Standards and Technology (USA), Standard Reference Material 999 (Potassium Chloride). Electrode used for measurement: Platinised Platinum Dip Cell. Reference: ASTM D-1125 Method A.

Accreditation:

Reagecon Diagnostics Ltd. is accredited to ISO 17025 by the American Association for Laboratory Accreditation, under scope 6739.03, for the test method, TPCOND, used to generate the above result. This accreditation deems Reagecon competent on a quality systems level and a technical level to perform the tests on the scope of accreditation. Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPCOND is technically valid and is supported by appropriate uncertainty measurements.

Date of Issue of the Certificate:

15/11/2024

QC Laboratory Techinician

gabriel

ONDIANO Gabriel

All raw materials used to prepare this product are of high purity.

*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL001