



Certificate of Analysis

Turbidimetric Calibration Standard

20 NTU Ratio Turbidity Standard

Product No:
Lot No:
Expiry Date:

CRSR-20-100
CRSR2020L1
28/11/2022

Certified Value:
Date of Measurement:
Assay Limits:
19 - 21 NTU

19 NTU
19/11/2020

Method:

Tested by Turbidity Measurement to In-House Method TPTURB.

Traceability:

The certified value is directly traceable to the Formazin Primary Standard for Turbidity, as adopted by NIST and other National Metrology Institutes.

Reference:

Water Quality-Determination of Turbidity - International Standard ISO 17027, DIN 38 404 and NF EN 27027.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

Date of Issue of the Certificate: 19/11/2020

A handwritten signature in blue ink, reading 'Patricia de Guzman'.

QC Technician Patricia de Guzman

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-20L19

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com