



Certificate of Analysis

Turbidimetric Calibration Standard

50 NTU Non Ratio Turbidity Standard

Product No: CRS-50-100
Lot No: CRS5021J1
*Expiry date *:* 28/09/2023

Certified Value: 51.8 NTU
Date of Measurement: 20/09/2021
Date of Sample Receipt: 20/09/2021
*Date of Manufacturing *:* 20/09/2021

* The detail above is based on information supplied in writing by Reagecon Manufacturing.

Assay Limits:

47.5 - 52.5 NTU

Method:

Tested by Turbidity Measurement to In-House Method TPTURB.

Traceability:

The certified value is directly traceable to the Formazin Primary Standard for Turbidity, as adopted by NIST and other National Metrology Institutes.

Reference:

Water Quality-Determination of Turbidity - International Standard ISO 17027, DIN 38 404 and NF EN 27027.

Date of Issue of the Certificate:

20/09/2021

QC Technician **Patricia de Guzman**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21J28

Tested by Reagecon Quality Control
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com