

Certificate of Analysis

European Pharmacopoeia

Primary Opalescent Suspension

Product No:	EPPOS01	Date of Measurement:	09/07/2025
Lot No:	EPPOS25G1	Date of Sample Receipt:	09/07/2025
Expiry Date*:	28/07/2026	Date of Manufacture*:	09/07/2025

Specification:	Mean Measured Value:
3800 - 4200 NTU	4003 NTU

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above.

Compliance:

This Product is manufactured according to the current edition of the European Pharmacopoeia.

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :
10/07/2025

QC Laboratory Technician
PILI Jashil

A handwritten signature in black ink, appearing to read "PILI Jashil".

All raw materials used to prepare this product are of high purity.

*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL002