

CERTIFICATE OF GRAVIMETRIC PREPARATION

PRODUCT: Calcium Standard (100 ppm Ca)
PRODUCT No.: 5000802C
MATRIX: H₂O + tr. Acetic Acid
LOT NO.: 5802C24H1
DATE OF PREPARATION: 19th August 2024
EXPIRY DATE: 28th August 2025
DENSITY VALUE: 1.001 g/ml @ 20°C

PREPARATION OF CONCENTRATE:

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

Analyte	Raw Material	Nominal mg/L	Actual mg/kg
Calcium, as Ca	Calcium Carbonate	1000	1000.2 ± 0.2 %

1000.2 mg/kg is equivalent to 1001 mg/L @ 20°C

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as ± 0.2 %. All values are verified by ICP-MS analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

DILUTION INSTRUCTIONS FOR PREPARATION OF Ph. Eur. 5000802 (100ppm Calcium Alcoholic)

1. To prepare Ph. Eur. 5000802 (100ppm Calcium Alcoholic) dilute this solution to 10 times it's volume with Alcohol R.
2. Prepare the dilute solution immediately before use.

TRACEABILITY IN THE PRODUCTION OF THIS STANDARD:

This product was prepared gravimetrically on a mass/mass basis, using a balance calibrated by Reagecon engineers with mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines. The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity, and linearity.

TEST METHOD:

The mean result of this standard was verified using a calibrated ICP-MS system according to an in-house test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high-performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By: QC Technician

A handwritten signature in dark ink, appearing to read 'gabriel', is written in a cursive style.

Date: 22nd August 2024

This certificate must not be reproduced except in full.