

CERTIFICATE OF GRAVIMETRIC PREPARATION

PRODUCT: IC Multi-Element Standard (6 elements)
PRODUCT No.: ICC-DX-611
MATRIX: H₂O
LOT NO.: ICCDX61124H1
DATE OF PREPARATION: 02nd August 2024
EXPIRY DATE: 28th August 2026
DENSITY VALUE: 1.002 g/ml @ 20°C

PREPARATION OF STANDARD:

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

Elements	Nominal mg/kg	Actual mg/kg	Actual mg/l @ 20°C
Calcium (as Ca ²⁺)	998	1004	1006
Ammonium (as NH ₄ ⁺)	399	401	402
Sodium (as Na ⁺)	200	201	201
Potassium (as K ⁺)	200	201	201
Magnesium (as Mg ²⁺)	200	201	201
Lithium (as Li ⁺)	49.9	50.2	50.3

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as $\pm 0.2\%$. All values are verified by ICP-MS analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

TRACEABILITY IN THE PRODUCTION OF THIS STANDARD:

This product was prepared gravimetrically on a mass/mass basis. The solute was weighed on a balance calibrated by Reagecon engineers using mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines. The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity, and linearity.

TEST METHOD:

The mean result of this standard was verified using a calibrated ICP-MS system according to an in-house test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high-performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By: QC Technician

A handwritten signature in cursive script, appearing to read 'gabriel', written in a dark ink.

Date: 13th August 2024

This certificate must not be reproduced except in full.