

Certificate of Analysis

Analytical Volumetric Solution

EDTA (disodium salt) 0.01M (0.02N)

Product No:

ED20011

Lot No:

ED200120K1

Date of Expiry Date:

28/10/2022

Mean Molarity:

0.01004 M @ 20°C ± 0.2%

Date of Measurement:

02/11/2020

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPEDTA. This volumetric solution was checked by means of Calcium Carbonate Analytical Volumetric Standard. It is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), 915b Calcium Carbonate. All raw materials used to prepare this product are of high purity.

This certificate relates solely to the lot number given above. The uncertainty of measurement has been calculated not to exceed ± 0.2% at 95% confidence level, i.e. coverage factor k =2.

Accreditation:

Reagecon Diagnostics Ltd. is accredited by the Irish National Accreditation Board, under scope 264T, for the test method, TPEDTA, used to generate the above result. This accreditation is intended only to certify that Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPEDTA is technically valid and is supported by appropriate uncertainty measurements.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

Date of issue of the certificate:

02/11/2020



QC Technician **Orla Sheehan**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-20L5

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com