

Certificate of Analysis

Brix Standard Stabilised

Sucrose (Brix) Standard 0°Brix @ 20°C

Product No:	BS00	Date of Measurement:	10/02/2023
Lot No:	BS0023B1	Date of Sample Receipt:	10/02/2023
Expiry Date*:	28/02/2024	Date of Manufacture*:	10/02/2023

Specification:	Mean Measured Value:
0.00	0.00 °Brix @ 20°C

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPRIA 01. The Brix value of the standard was determined using a high-performance, calibrated, temperature-controlled refractometer. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above. The uncertainty of measurement has been calculated not to exceed $\pm 0.11^\circ\text{Brix}$ (20°C) at 95% confidence level, i.e. coverage factor $k=2$.

Metrological Traceability:

This Test Result provides traceability to high purity ISO 17034 Certified Reference Materials.


Usage Instructions:

Prior to use, the contents of the dropper bottle should be brought to room temperature and mixed thoroughly by inversion to ensure a homogeneous mix.
Remove cap and discharge 4-5 droplets of liquid to waste to clear the nozzle.
Replace cap immediately after use.
Product must be stored at 2 - 8 °C when not in use.

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :
10/02/2023

Senior QC Technician
MALONE Eileen

A handwritten signature in black ink that reads "Eileen Malone".

All raw materials used to prepare this product are of high purity.
*The detail above is based on information supplied in writing by Reagecon Manufacturing.
Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing
This Certificate must not be reproduced except in full. Rev-QL1