



Shannon Free Zone,
Shannon, Co. Clare, Ireland
Tel: +353 61 472622
Fax: +353 61 472642
Email: sales@reagecon.ie
www.reagecon.com



CERTIFICATE OF CALIBRATION

PRODUCT: Density Standard 1.0265g/ml @ 25°C
PRODUCT NO: DEN25150
LOT NO: DEN21521A1
DATE OF TEST: 27 January 2021
EXPIRY DATE: 28 January 2023
SPECIFICATION: 1.0260 - 1.0270 g/ml @ 25°C
MEAN: 1.0264 g/ml @ 25°C

TEST METHOD:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high performance calibrated density meter according to TPDMA5000M. Density measurements are traceable to values for the density of pure water assigned by the PTB and NIST. The measurements were made in accordance with ASTM D4052 which is the standard test method for Density, Relative Density and API Gravity of Liquids by Digital Density Meter.

The uncertainty of measurement has been calculated not to exceed $\pm 0.16\%$ expanded uncertainty at 95% confidence level, i.e. coverage factor $k=2$.

ACCREDITATION:

Reagecon Diagnostics Ltd. is accredited by the Irish National Accreditation Board, under scope 264T, for the test method, TPDMA5000M, used to generate the above result. This accreditation is intended only to certify that Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPDMA5000M is technically valid and is supported by appropriate uncertainty measurements.

Approved By: QC Technician

Eileen Malone

Date: 28 January 2021

This certificate must not be reproduced except in full.