

Certificate of Conformity

Water and Synthetic Fresh Water Standards

Purified Water

Product No:
Lot No:
Expiry date :

H2O5
H2O25F2
28/06/2026

Date of measurement:

09/06/2025

Physical Description:

Clear, Colourless, Odourless Liquid
Formula: H₂O Molecular Weight: 18.02
CAS #: 7732-18-5

Specification:

Conforms to USP 36 for Packaged Purified Water
Max. specific conductance @25°C 1µS/cm-1(when first tested
pH: 5.0 - 7.0
Ammonium: Pass
Calcium: Pass
Carbon Dioxide: Pass
Chloride: Pass
Sulphate: Pass
Oxidisable Substances: Pass
Fluoride: <0.5ppm
TOC <500ppb

As per chapter 1231, this Purified Water meets the requirements of "bulk packaging for commercial use elsewhere". This product has not been sterilised. It is the users responsibility to assure fitness for use of this packaged article when used in manufacturing, clinical, or analytical applications where the pure bulk form of the water is indicated.

This is to certify that this product conforms to the specification given above when used as directed.

Date of issue of the certificate:

09/06/2025

QC Supervisor **Colin Ashton**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-25A1

Tested by Reagecon Quality Control
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com