

CERTIFICATE OF GRAVIMETRIC PREPARATION

PRODUCT: ICP Standard Yttrium 10,000 mg/L
PRODUCT No.: PY4B2
MATRIX: 2-5% HNO₃
LOT NO.: PY4221J1
DATE OF PREPARATION: 22nd September 2021
EXPIRY DATE: 28th September 2023
DENSITY VALUE: 1.042 g/ml @ 20°C

PREPARATION OF STANDARD:

All standard components have been pre-qualified/verified before use. All analytical measuring devices and instrumentation have been pre-calibrated. The actual concentrations reported below are based on this preparation methodology and compound impurities.

Raw Material	Purity %	Nominal mg/kg	Actual mg/kg
Yttrium Oxide	99.999	9597	9615 ± 0.2 %

9615 mg/kg is equivalent to 10019 mg/L @ 20°C

The expanded uncertainty (k=2) due to weighing, volumetric preparation and homogeneity is calculated in compliance with EURACHEM/CITAC Guide: Quantifying Uncertainty in Analytical Measurements as ± 0.2 %. All values are verified by ICP-MS analysis using externally sourced ISO 17034 accredited Certified Reference Materials as calibrants/quality controls where possible.

TRACEABILITY IN THE PRODUCTION OF THIS STANDARD

This product was prepared gravimetrically on a mass/mass basis. The solute was weighed on a balance calibrated by Reagecon engineers using mass standards traceable to the National and International primary standard of mass. Reagecon holds ISO 17025 accreditation for calibration of non-automatic weighing machines (265C). The resulting Balance Certificate of Calibration was issued in accordance with the requirements of ISO/IEC 17025. The balance was calibrated under monitored environmental conditions and atmospheric pressure. Tests were performed for capacity, readability, repeatability, eccentricity and linearity.

BALANCE ID No.: RRD077

CALIBRATION DATE OF BALANCE: 3rd August 2021

BALANCE ID No.: A-0944

CALIBRATION DATE OF BALANCE: 23rd July 2021

BALANCE ID No.: RRD078

CALIBRATION DATE OF BALANCE: 17th June 2021

CALIBRATION AUTHORITY OF BALANCE: Reagecon Diagnostics Ltd, ISO17025 Accreditation No. 265C.

TEST METHOD:

The mean result of this standard was verified using a calibrated ICP-MS system according to an in-house test method. The result reported in this certificate was confirmed by analysis of a sample of this lot taken at time of manufacture. The density of this standard was determined using a high performance calibrated density meter.

This certificate relates solely to the lot number given above.

Approved By: Paul O'Sullivan

Date: 30th September 2021

This certificate must not be reproduced except in full.