

Certificate of Analysis

Brix Standard Stabilised

Sucrose (Brix) Stabilised Standard 0°Brix @ 20°C

Product No:	BS00S	Date of Measurement:	30/04/2025
Lot No:	RIBS0025D1	Date of Sample Receipt:	30/04/2025
Expiry Date*:	28/04/2026	Date of Manufacture*:	30/04/2025

Specification:

0.00 °Brix @ 20°C

Mean Measured Value:

0.00 °Brix @ 20°C

Method:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. The Brix value of the standard was determined using a high-performance, calibrated, temperature-controlled refractometer. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above. The uncertainty of measurement has been calculated not to exceed $\pm 0.11^\circ\text{Brix}$ (20°C) at 95% confidence level, i.e. coverage factor $k=2$.

Metrological Traceability:

This Test Result provides traceability to high purity ISO 17034 Certified Reference Materials.

Composition:

This Product is Stabilised by the addition of 0.01% of preservative.

Usage Instructions:

Prior to use, the contents of the dropper bottle should be brought to room temperature and mixed thoroughly by inversion to ensure a homogeneous mix.

Remove cap and discharge 4-5 droplets of liquid to waste to clear the nozzle.

Replace cap immediately after use.

Product must be stored at 2 - 8 °C when not in use.

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :

30/04/2025

QC Supervisor

ASHTON Colin



All raw materials used to prepare this product are of high purity.

*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL006