

Shannon Free Zone

Shannon

Co. Clare

Ireland

Phone +353 61 472622

Fax +353 61 472642

Web www.reagecon.com

CERTIFICATE OF CONFORMITY

PRODUCT NAME: MELTING POINT STANDARD SET
BENZOPHENONE, BENZOIC ACID, ANTHRAQUINONE

PRODUCT No.: RMPSET2

LOT No. RMPSET224L1

EXPIRY DATE: 28/04/2027

This Melting Point Standard Set consists of the following Components;

Material No.	Material Product Description	Lot No
RMP048	Melting Point Standard Benzophenone 47 °C to 49 °C	R129324D1
RMP122	Melting Point Standard Benzoic Acid 121 °C to 123 °C	R01924E1
RMP284	Melting Point Standard Anthraquinone 283 °C to 286 °C	R129424G1

The Certificates of Analysis for the individual Melting Point Standard Set components are available under the above referenced Material No. and Lot No.

This certificate is electronically generated and does not require a signature.

Date: 09/12/2024

This certificate must not be reproduced except in full.