

**Shannon Free Zone** 

Shannon Co. Clare

Ireland

Phone +353 61 472622 Fax +353 61 472642 Web www.reagecon.com

## **CERTIFICATE OF CONFORMITY**

**PRODUCT NAME:** MELTING POINT STANDARD SET

SULPHANILAMIDE, CAFFEINE, VANILLIN

**PRODUCT No.:** RMPSET1

LOT No. RMPSET124M1

**EXPIRY DATE:** 28/06/2027

This Melting Point Standard Set consists of the following Components:

| Material No. | Material Product Description                           | Lot No    |
|--------------|--|-----------|
| RMP082       | Melting Point Standard Vanillin 81 °C to 83 °C         | R086224J1 |
| 1002         | Welting I one Standard Vanimin of C to 03 C            | 100022-31 |
| RMP165       | Melting Point Standard Sulphanilamide 164 °C to 166 °C | R086124F1 |
| RMP236       | Melting Point Standard Caffeine 235 °C to 238 °C       | R064824G1 |

The Certificates of Analysis for the individual Melting Point Standard Set components are available under the above referenced Material No. and Lot No.

This certificate is electronically generated and does not require a signature.

**Date:** 08/01/2025

This certificate must not be reproduced except in full