

Shannon Free Zone

Shannon

Co. Clare

Ireland

Phone +353 61 472622

Fax +353 61 472642

Web www.reagecon.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: MELTING POINT STANDARD ANTHRAQUINONE 283 to 286°C*

*This is a nominal range for product description purposes and is not the certified specification range.

PRODUCT No.: RMP284

LOT No.: R129422D1

DATE OF TEST: 22/04/2022

EXPIRY DATE: 28/04/2025

MENISCUS FORMATION: 283.6°C ± 0.3°C

COMPLETE LIQUEFACTION: 284.9°C ± 0.3°C

TRACEABILITY: The above certified temperature values are traceable to the ITS-90 Temperature Scale via instrument calibration using Certified Reference Materials provided by the UK Laboratory of the Government Chemist (LGC). The ITS-90 Temperature Scale is recognised as being the practical realisation of the SI Unit of temperature - the Kelvin by all leading international and national metrology institutes, including BIPM and NIST.

The above certified values were recorded at the first appearance of a liquid meniscus and the temperature at which no solid phase is apparent using an IA9200 melting point apparatus operating at a ramp rate of 1.0°C / min.

INSTRUCTIONS FOR USE:

Remove a sub-sample; lightly crush using a mortar and pestle to produce a fine powder, dry at 105°C for a minimum of 8 hours then store under reduced pressure in a desiccator containing a suitable desiccant.

This certificate is electronically generated and does not require a signature.

Date: 28/04/2022

This certificate must not be reproduced except in full.