

# Certificate of Analysis

## Total Organic Carbon Standard

### TOC Standard 1000mg/L C as KHP

Product No:	RTOCK18	Date of Measurement:	04/06/2025
Lot No:	TOCK1825F1	Date of Sample Receipt:	04/06/2025
Expiry Date*:	28/06/2026	Date of Manufacture*:	04/06/2025

**Specification:**

950 - 1050 mg/L C as KHP

**Mean Measured Value:**

990 mg/L C as KHP

**Method:**

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Tested in accordance with In-House Test Method. This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above.

**Traceability:**

This Test Result is traceable to the SI via National Reference Standards where applicable.

**Storage Conditions:**

Product must be stored at 2 to 8°C when not in use.

Products are manufactured under NSAI registered I.S EN ISO9001:2015 Quality Systems, registration no: 19.2769

Date of Issue of the Certificate :

04/06/2025

QC Laboratory Technician

PILI Jashil

A handwritten signature in black ink, appearing to read "PILI Jashil".

All raw materials used to prepare this product are of high purity.

\*The detail above is based on information supplied in writing by Reagecon Manufacturing.

Tested by Reagecon Quality Control Laboratories for Reagecon Manufacturing

This Certificate must not be reproduced except in full. Rev-QL1